

Purified Water Specification Testing Results (California Facility)

For the month of: February 2013 13032-13059

Listed below are the testing results for the purified water used in the manufacture of products at Hardy Diagnostics.

| Test | Testing Frequency | Units | In-House Specification (method detection limit) | Testing Data |
|---|--------------------------|---|--|---|
| Minimum Resistivity*1,3,4,5 | Continuous Monitoring | Megohm* cm | > 18.0 | 18.27 |
| pH* | Each use | N/A | 5.5 – 7.5 | 7.29 |
| Total Organic Carbon ^{1,3,5} | Monthly | ug/L | <500 | 49.4 |
| Heavy Metals* (Single) ^{1,3,4} (Cd, Cr, Cu, Ni, Pb and Zn) | Annually | mg/L | < 0.05 | Cd – ND Cr – 0.000092 Cu – 0.000050 Ni – ND Pb – ND Zn – 0.00083 |
| Heavy Metals* (Total)3,4 | Annually | mg/L | < 0.1 | 0.00097 |
| Ammonia/OrganicNitrogen ³ | Monthly | mg/L | < 0.1 | ND |
| Total Chlorine Residual ^{3,4} | Monthly | mg/L | <0.1 | ND |
| Maximum Bacterial Content ^{1,3,4,,5} | Weekly | colony forming units (CFU) per milliliter | <10 | <1 |
| Water Quality ^{3,4} ratio | Annually | ratio | 0.8 - 3.0 | 1.01 |
| Use Test (Student t)3 | Quarterly | N/A | <u><</u> 2.78 | 2.0 |
| Inhibitory Residue ⁴ | Annually | N/A | < 15% | 0.5% |
| Maximum Silicate, SiO ₂ 3,4 | Annually | mg/L | ≤ 0.05 | ND |

ND = Not Detected at or above the method detection limit.

References:

- 1. Preparation and Testing of Reagent Water in the Clinical Laboratory, C3-A4. Clinical Laboratory Standards (CLSI), Villanova, PA, 2006.
- 2. Quality Assurance for Commercially Prepared Microbiological Culture Media, M22-A3. Clinical Laboratory Standards Institute (CLSI formerly NCCLS), Villanova, PA, 2004.
- 3. Standard Methods for the Examination of Water and Wastewater, 21st ed., editors Andrew D. Eaton, Lenore S. Clesceri, et al. American Public Health Association, Washington, D.C. 2005.
- 4. Manual for the Certification of Laboratories Analyzing Drinking Water, Criteria and Procedures Quality Assurance 5th ed., Environmental Protection Agency (EPA).
- 5. USP. USP 35-NF 30, Water for Pharmaceutical Purposes <1231>. Rockville, MD: US Pharmacopeial Convention; 2012:886-907.

Wendy Hadley

Manager of Quality Control

Wendy Hadley

1430 W. McCoy Lane Santa Maria, CA 93455

(800) 266-2222 FAX (805) 346-2760 www.hardydiagnostics.com

^{*}Testing data is given as a monthly average.