

Purified Water Specification Testing Results (California Facility)

For the month of: February 2015 (15032-15059)

Listed below are the testing results for the purified water used in the manufacture of products at Hardy Diagnostics.

Test	Testing Frequency	Units	In-House Specification (method detection limit)	Testing Data
Minimum Resistivity*1,3,4,5	Continuous Monitoring	Megohm* cm	> 18.0	18.17
рН*	Each use	N/A	5.5 – 7.5	7.03
Total Organic Carbon ^{1,3,5}	Monthly	μg/L	<500	ND
Heavy Metals* (Single) ^{1,3,4} (Cd, Cr, Cu, Ni, Pb and Zn)	Annually	mg/L	< 0.05	Cd – ND Cr – ND Cu – 0.00036 Pb – ND Ni –0.000066 Zn – 0.00079
Heavy Metals* (Total) ^{3,4}	Annually	mg/L	< 0.1	0.0015
Ammonia/OrganicNitrogen ³	Monthly	mg/L	< 0.1	ND
Total Chlorine Residual ^{3,4}	Monthly	mg/L	<0.1	ND
Maximum Bacterial Content ^{1,3,4,,5}	Weekly	colony forming units (CFU) per milliliter	<10	<1.0
Water Quality ^{3,4} ratio	Annually	ratio	0.8 – 3.0	0.99
Use Test (Student t) ³	Quarterly	N/A	<u><</u> 2.78	<u><</u> 2.78
Inhibitory Residue ⁴	Annually	N/A	< 15%	-5.7%
Maximum Silicate, SiO ₂ 3,4	Annually	mg/L	<u><</u> 0.05	ND

ND = Not Detected at or above the method detection limit.

References:

- 1. Preparation and Testing of Reagent Water in the Clinical Laboratory, C3-A4. Clinical Laboratory Standards (CLSI), Villanova, PA, 2006.
- 2. Quality Assurance for Commercially Prepared Microbiological Culture Media, M22-A3. Clinical Laboratory Standards Institute (CLSI formerly NCCLS), Villanova, PA, 2004.
- 3. Standard Methods for the Examination of Water and Wastewater, 21st ed., editors Andrew D. Eaton, Lenore S. Clesceri, et al. American Public Health Association, Washington, D.C. 2005.
- 4. Manual for the Certification of Laboratories Analyzing Drinking Water, Criteria and Procedures Quality Assurance 5th ed., Environmental Protection Agency (EPA).
- 5. USP. USP 35-NF 30, Water for Pharmaceutical Purposes <1231>. Rockville, MD: US Pharmacopeial Convention; 2012:886-907.

Wendy Hadley

Manager of Quality Control

Ivendy Hadley

Hardy Diagnostics

1430 W. McCoy Lane Santa Maria, CA 93455

(800) 266-2222 FAX (805) 346-2760 www.hardydiagnostics.com

^{*}Testing data is given as a monthly average.