

Purified (Reagent) Water Specification Testing Results (Ohio Facility)

For the month of: August 2014 (H14213-H14243)

Listed below are the testing results for the purified water used in the manufacture of products at Hardy Diagnostics.

| Test | Testing Frequency | Units | In-House Specification (method detection limit) | Testing Data |
|---|--------------------------|---|--|--|
| Minimum Resistivity* ^{1,3,4,5} | Continuous Monitoring | Megohm cm | > 16.0 | 17.65 |
| pH* | Each use | N/A | 5.5 – 7.5 | 6.88 |
| Total Organic Carbon 1,3,5 | Monthly | µg/L | <500 | 172 |
| Heavy Metals (Single) ^{1,3,4} (Cd, Cr, Cu, Ni, Pb and Zn) | Annually | mg/L | < 0.05 | Cd – ND Cr – ND Cu – ND Ni – ND Pb – ND Zn – 0.0011 |
| Heavy Metals (Total) 3,4 | Annually | mg/L | < 0.1 | 0.0011 |
| Ammonia/Organic Nitrogen ³ | Monthly | mg/L | < 0.1 | ND |
| Total Chlorine Residual 3,4 | Monthly | mg/L | <0.1 | ND |
| Maximum Bacterial Content ^{1,3,4,5} | Weekly | colony forming units (CFU) per milliliter | <10 | <1 |
| Water Quality ^{3,4} ratio | Annually | ratio | 0.8 – 3.0 | 1.14 |
| Use Test (Student t) ³ | Quarterly | N/A | <u><</u> 2.78 | 2.0 |
| Inhibitory Residue ⁴ | Annually | N/A | < 15% | 5.3 |
| Maximum Silicate, SiO ₂ ^{3,4} | Annually | mg/L | <u><</u> 0.05 | ND |

ND = Not Detected at or above the method detection limit.

*Testing data is given as a monthly average.

References:

1. Preparation and Testing of Reagent Water in the Clinical Laboratory, C3-A4. Clinical Laboratory Standards (CLSI), Villanova, PA, 2006.

2. Quality Assurance for Commercially Prepared Microbiological Culture Media, M22-A3. Clinical Laboratory Standards Institute (CLSI – formerly NCCLS), Villanova, PA, 2004.

3. Standard Methods for the Examination of Water and Wastewater, 21st ed., editors Andrew D. Eaton, Lenore S. Clesceri, et al. American Public Health Association, Washington, D.C. 2005.

4. Manual for the Certification of Laboratories Analyzing Drinking Water, Criteria and Procedures Quality Assurance 5th ed., Environmental Protection Agency (EPA).

5. USP. USP 35-NF 30, Water for Pharmaceutical Purposes <1231>. Rockville, MD: US Pharmacopeial Convention; 2012:886-907.

Wendy Hadley

Wendy Hadley Manager of Quality Control

Hardy Diagnostics 1430 W. McCoy Lane Santa Maria, CA 93455 (800) 266-2222 FAX (805) 346-2760 www.hardydiagnostics.com